MALLORY Mallor	y Sonalert Pr	oducts,Inc.	Part #: SB1	T12MMEPC
Sales Outline Drawing		Revision:	D	
Specifications:	cUI Recogniz	ed to IEC60601-1-8 -	Fauinment - Supply	Failure Melody
Sound Level Category:	Silence	Loud	Loud	Loud
Mode of Operation:	Silence - NO SOUND	Low Priority Alarm 2 beeps repeating every 30 sec	Medium Priority Alarm 3 beeps repeating every 7.5 sec	High Priority Alarm 10 beeps repeating every 2.5 sec
Loudness @ 10cm:		85 to 95 dB(A) Typ.	85 to 95 dB(A) Typ. 85 to 95 dB(A) Typ.	
Frequency:		150 to 4000 Hz		
Connection:	Connect "+" to "+" & "-" to "-" of Power Supply		Connect "+" to "+" & "-" & M to "-" of Power Supply	I X. H to "-" of Power
Voltage Rating:		9 to 1	12 VDC	
Average Current Draw:	<15mA Typ.	<15mA Typ.	<15mA Typ.	<20mA Typ.
Ipk Current*:	75mA			
Housing Material:	ABS 777D Plastic Resin, Black			
Storage Temperature:	-40° to +85° C			
Operating Temperature:	-30° to +65° C			
Weight Typical:	0.6 oz (17g)			
Options:	Please contact factory.			
Dimensions: Inches [m	-	cUL Recognized		ROHS Compliant
	PATEN	IT NUMBER: 7880593		
Before Code Date: 154	6 Not cUL Recognized	ot cUL Recognized After Code Date: 1547		cUL Recognized
	Ø1.750 [Ø	Ø44.5]	0.334 [8.5]	
	0.100 [2	0.200 [5.1		MALE AY MAIDRY Medical

*Ipk = Instantaneous Peak

0.551 [14.0]

Ø0.039 [Ø1.0] 5 Holes Req'd

0.229 [5.8]